

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 11:22

Date of Inspection: 05/12/2010

Serial Number: 80-000907
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:200912C Exp: 12/01/2011	0.08g/210L Test (g/210L) Lot#:200811C Exp: 11/10/2010	0.20g/210L Test (g/210L) Lot#:200908C Exp: 08/27/2011	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:831802I Exp: 11/22/2010
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.047	0.078	0.195	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DENNIS J MALLARDI

Signature and Printed Name

05/12/2010
Date