

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 11:08

Date of Inspection: 05/12/2010

Serial Number: 30-000912  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 200912C Exp: 12/01/2011	0.08g/210L Test (g/210L) Lot#: 200811C Exp: 11/10/2010	0.20g/210L Test (g/210L) Lot#: 200908C Exp: 08/27/2011	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 831802I Exp: 11/22/2010
0.000	0.048	0.079	0.195	0.079
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.079	0.195	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DENNIS J MALLARDI

Signature and Printed Name

05/12/2010  
Date